

DTS-180 VET Dental System

Operation Manual





Intellectual Property Right

The intellectual property rights of this product and its instruction manual belong to TOOTOO MEDITECH CO.,LTD.(hereinafter referred to as the MEDITECH), including,but not limited to patent ,trademark,copyright etc. MEDITECH reserves final interpretation right of this instruction manual.

MEDITECH have the right to use the instruction as confidential information.Any individual and/or organization shall not disclose the instruction of all or part of the information by any means without MEDITECH 's written permission.Nor shall any other person or organization be allowed to obtain all or part of the information of this instruction manual by any means.

No individual and/or organization shall publish, modify, reproduce, issue, rental, adapted, and translated into other languages without MEDITECH 's written permission.

Statement

MEDITECH reserves the right to modify the content of this manual without prior notice.

MEDITECH reserves the right to change the technology without prior notice.

MEDITECH reserves the right to modify the product specifications without prior notice.

MEDITECH does not guarantee the information in any forms,including (but not limited to) the responsibility of proposing the implied merchantability and suitability for a specific purpose.

MEDITECH in only the following conditions is considered to be responsible for the safety, reliability and performance of the instruments,i.e.

Assembly operation,expansion,adjustment,improvements and repairs carried out by MEDITECH authorized personnel.

Relevant electrical equipment in line with national standards.The instrument is operated according to the instruction manual.

MEDITECH is not responsible for the products' safety, reliability or operation status in the following conditions:

Components are disassembled, stretched or debugged



Non-authorized personnel by MEDITECH repairs or alters to the products without following the manual.



CONTENT

DTS-180 VET Dental System	1
Operation Manual	1
Intellectual Property Right	2
Statement	2
1-Introduction	5
1.1 Application	6
1.2 Note	6
1.3 Products List	7
1.4 Performance parameter	8
1.5 Environment requirements for equipment	9
2-Getting Started	9
3-Product Structure and Guide	10
3.1 Water Filtration System	15
Refilling Water Bottles	15
3.2 Way Air/Water Syringe	16
Tips Removal	16
3.3 High Speed Handpiece	16
Bur Replacement	17
3.4 Low Speed Handpiece	17
3.5 Ultrasonic Scaler Handpiece	18
3.6 Handpiece Pressure Regulation	19
3.7 Silent Oil-Free Air Compressor	21
3.8 Dental Curing Light (Optional)	21
4-System Safety	21
4.1 Attentions	23
4.2 Precautions	23
5-Maintenance	24
5.1 Handpiece Maintenance & Sterilization	24
5.2 Air Compressor Maintenance	25
6-Troubleshooting	26
6.1 High Speed Handpiece	26
6.2 Low Speed Handpiece	27
6.3 Way Air/Water/Spray Syringe	28
6.4 Air Compressor	28
7-Warranty	29



1-Introduction

Sincerely thanks for selecting the Veterinary Dental System manufactured by MEDITECH .

MEDITECH Dental Systems are designed with unique and practical features to make your work even more satisfying and productive.

Video Instructions

As for how to assemble and how to maintain MEDITECH dental system. All instructions are conveniently delivered to you by video instruction step by step.

Control Top

The top is a non-porous work surface which is easy to sanitize. It provides a detachable rubber plate with magnetism around the edge to supply a safe area to store the tools and prevent burs falling on the floor during procedures.

Silent Oil-Free Air Compressor

MEDITECH takes an Oil free compressor -low maintenance, oil free air. Mounts underneath the unit into the body, with coolant system and noise reduction system, supplying air wherever the unit is to be used. It is more convenient than a noisy compressor which is usually situated in another room and requires a length of airline.

Reliable Air Lines and Pipeline

All inter airlines are bundled with brass air tip in case loosen after longtime using. Outer airlines are locked with stainless connectors. And using the rubber pipeline instead of PCV pipeline to against the aging and crack.

Air Particle Filters

Air particle filters are installed on MEDITECH dental systems to help reduce particles entering the air supply and reduce the potential damage to handpiece, also reduce the unpleasant environmental odors.

LED Advantage



The self-generated LED provides daylight-quality, neutrally-coloured white light at 6,000 Lux. LED's offer a robust, long lasting Light source that barely heat up, are shock resistant and that do not emit any harmful IR and UVA rays. DTS-180 is designed LED advantage into high speed handpiece and ultrasonic scaler.

Please read this instruction manual and all other auxiliary materials carefully before installing and using the product, it will be helpful to work with it better.

1.1 Application

Veterinary dental system is a dental platform that integrates polishing, repairing, cutting and auxiliary root canal treatment. It is used for dental beauty, tooth removal, oral tumor cutting, etc., providing a convenient and favorable operating environment for animal dentists.

1.2 Note

MEDITECH dental system should be operated by the qualified trained person

— for Veterinary use only

For operator safety, it is recommended that safety goggles or visor, face mask and surgical gloves be worn during all procedures.



1.3 Products List

Name	Amount	Standard/Optional	Remark
Dental system host	1 set	Standard	Power cord Foot switch Detachable magnetic rubber tray
3 way air/water syringe	1 pcs	Standard	2pcs of syringe tips
High speed handpiece	1 pcs	Standard	with LED light
Low speed handpiece	1 pcs	Standard	Straight type/Contra-angle type
Ultrasonic Scaler handpiece	1 pcs	Optional	with LED light
Suction handpiece	1 set	Optional	With tube, take liquid out from pet's mouth
Dental Curing Light	1 pcs	Optional	Promote the hardening of resin materials
Clean water bottle	1 pcs	Standard	
Waster bottle	1 pcs	Optional	
Scaler tip	1 package	Optional	4pcs
Lubricant for 1pcs handpieces	1 bottle	Optional	30ml~500ml



1.4 Performance parameter

Parameter	Description
High Speed Handpiece	Working Pressure:0.3-0.35Mpa
	Rated Speed:320,000-400,000RPM
Low Speed Handpiece	Working Pressure:0.3-0.35Mpa
	Rated Speed:18,000-20,000RPM
Ultrasonic Scaler Handpiece	Output Power: 3-20W
	Water Pressure: 0.01-0.5Mpa
Air Compressor	Power: 550W 120/220V, 50/60HZ
	Current: 2.4A
	Airflow: 70L/min
	Output Pressure:0.8Mpa
Noise	≤60db
Size(L*W*H):	550*520*900mm
Net Weigh	40KG




1.5 Environment requirements for equipment

Parameter	Description
Operation Environment	Temperature:5-40°C
	Humidity:30%-85%(non-condensing)
	Air pressure:70-106kPa
Storage Environment	Temperature:-40°C~55°C
	Humidity:<85%(non-condensing)
	Air pressure:70kPa~106kPa
Working Power	120/220V,50/60HZ

2-Getting Started

- 1.When you take the machine away from the bottom plate,do not take the Control Top,but take the Body, in case damage the Control Top.
- 2.Loosen the four castor wheels under the body to move machine or lock them to fix it unmovable.
- 3.Unscrew the clean water supply clear bottle,then push out the accessory box,which includes all handpieces and pedal inside. Take out of all accessories,assembly and put them in holder correspondingly,and place the foot switch on the floor in front of machine.
- 4.Fill the clear plastic water bottle with distilled water,not too full,and screw the bottle into the hole with pipe under the Control Top, tight the bottle.
- 5.Place the 3 way air/water/spray syringe in its handpiece holder.
- 6.Assemble the handpieces according to the manufacturers instructions. Screw the gray handpiece air/waterline onto the handpiece and place in correctly labeled handpiece holder on the dental system.



7. Connect the host to power supply with the supplied power cord.
8. Turn on the power and the compressor start to pressurize. It needs 1 minutes to get ready on working pressure at the first time by using.
9. MEDITECH factory sets the working pressure. You can use it directly. But you still can adjust individual handpiece pressure by corresponding knob easily.

If adjustment is required, make sure your compressor is set to deliver the correct working pressure. When you depress the foot switch, the air pressure regulator gauge should be within working pressure.

10. Check handpiece and ultrasonic scaler operation:

- a) Remove High Speed Handpiece from holder and depress foot switch, then handpiece starts to operate.
- b) Replace the High Speed handpiece and remove the Low Speed Handpiece from holder and depress foot switch, then handpiece starts to operate.
- c) Replace the Low Speed handpiece and remove the Ultrasonic Scaler Handpiece from holder and depress foot switch, then handpiece should start to operate.


NOTE: The DTS is fitted with automatic handpiece holders. Only one handpiece (HS or LS) can be used at any one time. Make sure that you put the HS/LS handpiece back to the holder before picking up LS/HS handpiece.

11. Check water and air supply to the 3 way air water syringe. Depress the right hand button for air and left hand button for water or both together for mist.
12. Pressurize water system by moving the water system switch to ON located on the control panel.

NOTE: When filling distilled water bottles de-pressurize the water system, moving the system switch to OFF.


13. Water volume. Remove HS/LS/Ultrasonic Scaler Handpiece from holder and depress foot control pedal. Adjust water spray volume by flow control knob, and then, if required, to get a fine mist by rotating vaporization control knobs.
14. Adjust Ultrasonic Scaler Handpiece by the white color knob, located on the last of control panel.
15. HS and LS handpieces should be lubricated regularly to get optimal performance and life. Recommend lubricated before using each time. The foot pedal should be depressed for 15 seconds after oiling to ensure the lubricant will reach the turbines and working portions of the handpiece.

NOTE: The lubricant is placed in the smaller of the 2 large holes. "Air entry hole" at the base of the HS/L S handpiece. Using before and after using for 30 minutes, please drop with lubricant.



3-Product Structure and Guide

Front



- 1) Suction handpiece
- 2) 3 way air/water/spray syringe
- 3) High speed handpiece
- 4) Waste bottle
- 5) Caster
- 6) Rubber tray for instruments
- 7) Handpiece pressure gauge
- 8) Clean water
- 9) Ultrasonic scaler handpiece
- 10) Low speed handpiece
- 11) Foot switch
- 12) Dental Curing Light



Knob of Water Regulate: to adjust the water volume

Knob of Air Regulate: to adjust the air volume

Knob of Scaler: to adjust the scaler power

Switch of Water ON/OFF: toggle switch to take on/off watering



Adjust the air pressure of handpiece 3-high speed handpiece

Adjust the air pressure of handpiece 2-low speed handpiece

Adjust the air pressure of handpiece 1-scaler handpiece

Back of Dental System



1: connect to air compressor

2: main power supply cable

Back of air compressor



AIR OUTLET: connect to dental system

SWITCH: power switch ON/OFF

AC INPUT: to power supply



3.1 Water Filtration System

MEDITECH dental system is supplied with a clear and a white water bottle. The clear bottle should be filled only with distilled water to assist in preventing mineral or bacteria build up in the handpieces. This water bottle supplies the handpieces, 3 way air water syringe and ultrasonic scaler.

The water filtration system has been preset, and can turn on and off it by toggle switch to get the water supply.

The white waste water bottle to collect the liquid took out from pet's mouth by suction handpiece and tube.

Refilling Water Bottles

To refill the distilled water, toggle the pressure switch on the dental system to OFF -depressurizing the system. After the system has been depressurized, unscrew the bottle and refill with either distilled water. After filling the water bottle screw it back into the holder and toggle the pressure switch to ON, repressurize the system.

DO NOT OVERTIGHTEN THE BOTTLES.

Please Note: When the day's dental procedures are complete, toggle the pressure ON/OFF switch to the OFF position, depressurizing the distilled water and waste water bottles.

NOTE: Failure to use distilled water will void the machine warranty, which will result in crystal deposits forming, damage handpieces gear/motor and block valves.

Filtration System

The use of distilled water is mandatory in the water bottles, as local water may clog the system and can corrode the brass valves in the system, apart from the risk of a bacterial build up.

Chlorine in tap water can cause flocculation, clogging filters and valves. If you have no choice to use tap water, you need buy Inline filter, and fix it on the pipeline in the supply bottles. This kind of filtration help prevent particles from entering the system and blocking handpieces. This kind of filter is located in the distilled water bottle.

Attention: This kind of inline filter need to be checked and replaced if they appear dirty.



Cleaning

Bottles may be sterilized with ethylene oxide or by chemical sterilization

3.23 Way Air/Water Syringe

Your choice of air, water, or fine mist spray. Depress the right-hand button for air (with air injection icon) and the left-hand button for water (with water drop icon) and both for a fine mist.

User could select water or disinfectant by the water/disinfectant toggle switch.

The 3 way syringe is ideal for flushing away debris during prophylaxis or drying areas during endodontic procedures.



- 1) Out ring housing
- 2) Air
- 3) Water

Tips Removal

The tip of the 3 way syringe may be removed for cleaning and autoclaving.

Simply depress the outer ring housing and remove the tip. (See picture as above)

3.3 High Speed Handpiece

The High Speed Handpiece is designed to operate well at pressure gauge in Green area (Yellow means not enough, Red means too high pressure).

The standard HS handpiece is designed at speeds of up to 400,000 rpm. A very light touch is necessary as handpieces operate with very low torque and high



speed. Pressure on the handpiece of 30gms(1ounce) or more, will cause the bur to stall.

Distilled water can be ejected by HS handpiece. Fine mist is easily adjusted by knob on control panel.

Note: All handpieces should be lubricated before use.

Bur Replacement

1. Depress the head back of handpiece with your thumb and pull out the bur.
2. Insert the bur into the hole all the way until it stops, and then depress the head back of handpiece with your thumb, at the same time, depress the bur again until it stops.
3. To change the Bur repeat step 1 and step 2.

To increase the life of the Handpiece, oil between each use and use a new Bur between each patient pet. See Handpiece manual.



3.4 Low Speed Handpiece

The low speed handpiece kinds include Straight type without coolant and Contra-angle type with water pipeline.



Both of them can adjust the air pressure and can adjust the speed and rotation direction by the knob located on handpiece itself. The straight type reducing head is designed for polishing. The speed of the polishing cup should be maintained at less than 5,000 r.p.m. The reducing attachment therefore allows the motor speed and torque to be maintained while reducing the polishing speed, not generate excessive heat. But the Contra-angel type can get more position and application than Straight type, also it can coolant by distilled water from pipeline.

Please refer to the video to understand how to install the polishing head (lock and unlock), adjust speed and direction.

3.5 Ultrasonic Scaler Handpiece

Find the red dot to match and assemble the ultrasonic scaler handpiece.

Using the plastic wrench to lock and unlock the tips. It allows users to set power output at their desired level of operation by white knob which is located on control panel.

The ultrasonic scaler with LED light technology has been designed for use in the areas of scaling, periodontology, endodontics, and micro-preparation. Lightweight handpiece with foot switch ON/OFF control.



3.6 Handpiece Pressure Regulation

The high and low speed handpieces should be set to a specific pressure. Even factory sets the default working pressure, user still can adjust it. The handpiece pressure adjustment is located on the front control panel, one is for HS, one is for LS, and one is for scaler. See below picture and notes.



Adjust the air pressure of handpiece 3-high speed handpiece

Adjust the air pressure of handpiece 2-low speed handpiece

Adjust the air pressure of handpiece 1-scaler handpiece



Knob of Water Regulate: to adjust the water volume

Knob of Air Regulate: to adjust the air volume

Knob of Scaler: to adjust the scaler power

Switch of Water ON/OFF: toggle switch to take on/off watering

Pressure Regulation Method:

1. Turn the unit on and remove the HS handpiece to be adjusted.
2. Depress the foot switch and turn the knob clockwise or anti-clockwise, until the desired pressure is displayed on the pressure gauge.
3. Make sure that the HS handpiece is plugged into its bracket, then take off the LS handpiece and then adjust the corresponding knob, until the desired pressure is displayed on the pressure gauge.

KEEP IN SAFE PRESSURE VALUE

pressure
gauge



3.7 Silent Oil-Free Air Compressor

The compressor is perhaps the most crucial part of any dental system as it supplies the valuable air needed to run handpieces and pressurize the water and cooling system.

Parameters preset before shipment, no any change if there is no special requirement. **Just be a attention: Connect The Host to The Appropriate Voltage.**

High voltage used on this equipment.

Isolate from mains before removing any covers, which should only be carried out by a qualified electrician.

Ensure power outlet is adequately rated for this unit.

Isolate power supply and relieve air pressure prior to service tasks.

This appliance is not intended for use by persons (including children) with reduced physical, sensory or mental capabilities, or lack of experience and knowledge, unless they have been given supervision or instruction concerning use of the appliance by a person responsible for their safety.

3.8 Dental Curing Light (Optional)

Dental curing light for filling material refers to LED light curing lamp. Its main functions include promoting the hardening of resin materials, assisting in judging whether the filling is tight, and preventing secondary caries. Generally, LED light sources are used, which can convert blue light of a specific wavelength into ultraviolet light, causing the photosensitizer to produce a cross-linking reaction to achieve the repair effect.

1. Promote the hardening of resin materials: When filling teeth, it is usually necessary to use resin materials to completely fill the defective area and wait for a certain period of time before it can be completely solidified. By using a light curing lamp, the resin material can react faster to achieve the hardening effect.

2. Assist in judging whether the filling is tight: If the patient's tooth filling falls off, it may be caused by the lack of adhesion between the filling and the tooth. At this time, the doctor can observe whether the filling has been fully adhered to the tooth by using a light curing lamp, so as to determine whether the filling is not tight.



Correct maintenance method

1. Clean the lamp head: When the dental curing light lamp is working, the lamp head will come into contact with the patient's mouth and is easily contaminated. Therefore, you need to wipe the lamp head with an alcohol cotton ball or clean gauze before each use.
2. Prevent inversion: Keep the curing light in a stable position and do not invert it. Inversion can easily cause air to enter, affecting the effect of the light source.
3. Regular cleaning: During daily use, the dental curing light will be contaminated with a variety of bacteria and viruses and needs to be cleaned regularly.
4. Reasonable use time: The service life of LED light source lamp beads is very long, but it is also necessary to control the reasonable use time to reduce the aging problem of the lamp beads.
5. Maintain ventilation: When using a curing light, maintain good ventilation. Do not point the lamp head directly at the patient's mouth to avoid generating odor.

4-System Safety

Please read safety instructions carefully. For the sake of safety, please be sure to be aware of the followings:

4.1 Attentions

1. Correct cable connection
Ensure the correct voltage and securely grounding.
2. Avoid exposure of all electronic cables
Do not touch any electronic devices and circuits inside the device
3. Please stop the device in case of suspected failure
If there is potential danger in the device or it cannot be operated properly, contact authorized technical personnel.
4. Correct device (all handpiece) connection
Connect device correctly to avoid difficulties in operation due to wrong connection, resulting in disconnection.

4.2 Precautions

1. Fasten power cord to prevent poor power contact.
2. Pay attention to shock, water, humidity, pressure, and fire-proof.
3. Pay attention to handling strength and position during moving and handling, to prevent damage or dump the instrument.
4. Untrained personnel are prohibited from operating the equipment. Please use a reliable grounding device.
5. Users shall undertake corresponding duties due to any improper cleaning, maintenance and operation.
6. In case that the equipment host is disassembled without permission of MEDITECH, it shall no longer be guaranteed by the MEDITECH with technical maintenance service. Please contact authorized personnel or MEDITECH for support in case of any technical problems.

5-Maintenance

5.1 Handpiece Maintenance & Sterilization

All handpieces have a working life but it can be extended by maintaining with the correct operating pressures and speeds as well as adequate lubrication.

Device	Maintenance		Sterilization	
	Daily	Yearly	Before	After
High Speed Handpiece	Oil before or after use Oil in smaller of 2 big holes Depress the foot pedal for 10 seconds to ensure the oil reaches the Turbine	Replace turbine and o'rings and sealing gaske if worn	1. Thoroughly cleaned to remove any foreign particles, eg ;blood, saliva etc 2. Clean the exterior of the handpiece with any good surgical disinfectant. Be sure to rinse and dry thoroughly 3. Remove moisture with a towel or use air from the 3 way air water syringe 4. Take off the bur and lubricate handpiece turbine with lubricant 5. Seal the handpiece in sterilization tubing bags before autoclaving	1. Allow handpieces to cool down then lubricate handpiece thoroughly 2. Stored in another container to avoid recontaminating
Low Speed Handpiece	Same as above (HS instruction)	Replace turbine and 'o'rings and sealing gasket if worn	Almost same as HS, but need to remove head from handpiece, and oil through both end openings After lubricating items, assemble and operate for a few seconds, or use air from the 3 way air water syringe to expel excess lubricant	Same as above (HS instruction)
Ultrasonic Scaler Handpiece	1. Clean with any good surgical disinfectant 2. Be sure to rinse and dry thoroughly	Replace turbine and o'rings and sealing gasket if worn	1. Clean and remove the 'o' ring and led diffuser 2. Clean with any good surgical disinfectant, but be sure to rinse and dry thoroughly.	Stored in another container to avoid recontaminating.



The working life of most handpieces turbines is 12-24 months, however, it will depend upon the degree of care and the frequency of use.

Note: Make sure that the turbine cartridge and head cavity are cleaned thoroughly before autoclaving.

If autoclaved with bloody particles in the turbine housing, autoclaving will cause particles to be scorched and damage to the turbine. Prior to autoclaving, lubricate items.

**Note: Do not autoclave the handpiece with a bur.
Do not exceed temperatures of 273°F or 134°**

Lubricant after cooling down, do not allow hand-piece to sit in sealing bag overnight or for a long period.

5.2 Air Compressor Maintenance

Turn off the power before **Daily Maintenance**

Yearly Maintenance:

It is recommended to replace the inlet filters and pressure regulator filters every year, please contact the after-sale service personnel for dealing with it.



6-Troubleshooting

The chapter describes problems (faults) and suggested measures that are common in using the product

6.1 High Speed Handpiece

Problem	Caused	Solution
No Water Flow in Handpiece	1. Compressor not ON	Turn compressor ON and check that ON/ OFF tap is in the ON position
	2. Water flow control not ON	Turn water control to ON
	3. Clean Water bottle empty	Fill water Bottle
	4. The water line is blocked	Check lines and pressure.
	5. Check if 3 way syringe has water	If yes, lack of water must be in the valves or the handpiece. The 3 way syringe is the quickest way to check for water
	6. Use of tap water, blocking water filter	Valves will need replacing.
	7. Pressure regulator is loose	Contact after-sales personnel.
Mineral Accumulation	1. Use of tap water-mineral or chemical action blockage will usually take 3-4 years to occur.	Pass a fine wire through the water intake hole. Pass another fine wire through the hole in the handpiece head. Pass the wire as far as possible each way, blow out the loosened mineral accumulation
	2. Verdigrise build up in brass water valves due to chemical reaction with tap water	Replace valves
Use of Chlorhexidine	1. Most of disinfectants will cause solution deposit and block valves Note: insufficient dilution can also cause deposition.	Do not dilute beyond its label recommendations Contact the after-sales personnel if request
	2. Chlorine in tap water will cause a flocculation and block filters and valves	Use distilled water only
Handpiece won't run	Drive head damaged	Contact the after-sales personnel
LED is not on	LED damaged	Contact the after-sales personnel



6.2 Low Speed Handpiece

Problem	Caused	Solution
No Water Flow in Handpiece	1. Compressor not ON	Turn compressor ON and check that ON/OFF tap is in the ON position
	2. Water flow control not ON	Turn water control to ON
	3. Clean Water bottle empty	Fill water Bottle
	4. The water line is blocked	Check lines and pressure.
	5. Check if 3 way syring has water	If yes,lack of water must be in the valves or the handpiece.The 3 way syringe is the quickest way to check for water
	6. Use of tap water,blocking water filter	Valves will need replacing.
	7. Pressure regulator is loose	Contact after-sales personnel.
Mineral Accumulation	1. Use of tap water-mineral or chemical action blockage will usually take 3-4 years to occur.	Pass a fine wire through the water intake hole. Pass another fine wire through the hole in the handpiece head. Pass the wire as far as possible each way, blow out the loosened mineral accumulation
	2. Verdigrease build up in brass water valves due to chemical reaction with tap water	Replace valves
Handpiece won't run	1. No or low air pressure	Check the air line to be sure the air is turned on and the line is unobstructed
	2. Inferior oil, too much or too little oil or water in the turbine	Refer to lubrication instructions
	3. Drive head damaged	Contact the after-sales personnel
The polishing head has no effect	Polishing head is loose	Reinstall the polishing head and screw it.
Overheat of polishing head	Pressure is too large	Adjust the speed of handpiece
The polishing head not run	Handpiece doesn't run	Check the manual of LS handpiece, depress the foot switch, adjust the speed control knob



6.33 Way Air/Water/Spray Syringe

Problem	Cause	Solution
No Water Flow in Handpiece	1)Compressor not ON	Turn compressor ON and check that ON OFF tap is in the ON position.
	2)Water flow control not ON	Turn water control to ON
	3)Water/Disinfectant bottle empty	Fill water/Disinfectant Bottle.
	4)The water/air line is blocked.	Check lines and pressure.
	5)Use of tap water,blocking water filter	Valves will need replacing
	6)Pressure regulator is loose.	Contact after-sales personnel.
Mineral Accumulation	1)Use of tap water -mineral or chemical action blockage will usually take 3-4 years to occur	Pass a fine wire through the water intake hole.Pass another fine wire through the hole in the handpiece head.Pass the wire as far as possible each way,Blow out the loosened mineral accumulation
	2)Verdigrease build up in brass water valves due to chemical reaction with tap water	Replace valves.

6.4 Air Compressor

Check the supply voltage and grounding.

Check whether the power cord in socket is well plugged.

Contact the after-sales personnel.

7-Warranty

The warranty of this equipment starts from the date of leaving the factory. During the warranty period, the equipment cannot be used normally due to problems such as materials and process defects. MEDITECH undertakes after-sales service such as equipment maintenance and parts replacement.

Any damage caused by improper use or over-range use is not covered by the warranty. If repair or replacement of parts is required, the cost will be borne by the user.

If there worked equipment was found to have been unauthorized disassembly, MEDITECH will not provide after-sales service such as warranty, free maintenance and parts replacement.

The warranty statement (including its restrictions) is exclusively issued by MEDITECH and covers all other warranties.



TOOTOO MEDITECH CO., LTD.

Address: 3A, Phase II, Smart Home, Baolong Street,
Longgang District, Shenzhen, Guangdong, China
Tel: +86-755-28306385
Email: info@newtootoo.com
Web: www.newtootoo.com